


CERTIFICATE OF ANALYSIS

Product information

Name	BCR(E14)-ABL1(E2)_P210 Fusion
Catalog No.	CBP20031R
Lot No.	CBFA23091702
Format	RNA
Left Gene	BCR (E14)
Left Breakpoint	chr22:23632600:+
Right Gene	ABL1 (E2)
Right Breakpoint	chr9:133729451:+
Ploidy	2.96
Buffer	RNase-free H ₂ O
Storage Conditions	-80°C
Date of Manufacture	2023/09/17
Expiry	12 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	60.00 ng/uL±10%	58.90 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.93 by Nanodrop detection	√
RQN value	>9	9.7 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing	√
Copy number	Positive	763.21 copies/ng with bio-rad on-step Kit detection	

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****