



CERTIFICATE OF ANALYSIS

BCR-ABL1 Fusion

Product information

Name	BCR-ABL1 Fusion
Catalog No.	CBP20031R
Lot No.	CGD2021071303
Format	RNA
Left Gene	BCR
Left Breakpoint	chr22:23632600:+
Right Gene	ABL1
Right Breakpoint	chr9:133729451:+
Ploidy	2.96
Buffer	RNase-free H2O
Storage Conditions	-80°C
Date of Manufacture	Jul.-13-2021
Expiry	18 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	5ug/vial*2vials	5ug/vial*2vials	√
Concentration	Qubit 3.0	60 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.1	1.91	√
RNA electrophoresis	No Degradation	 <p>M: Marker 1: BCR-ABL1 Fusion</p>	√
Genotype	Sanger Sequencing		√

Caution:


No.222 Changhong Road, Yuhua
 District, Nanjing, Jiangsu
www.cobioer.com

Tel: 4008-750-250
 QQ: 4008-750-250
 Email: sales@cobioer.com

For research use only.

*the expiry is tested in house.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****