


## CERTIFICATE OF ANALYSIS

### Product information

Name	BCR(E14)-ABL1(E2)_P210 Fusion
Catalog No.	CBP20031R
Lot No.	CGFA23042703
Format	RNA
Left Gene	BCR (E14)
Left Breakpoint	chr22:23632600:+
Right Gene	ABL1 (E2)
Right Breakpoint	chr9:133729451:+
Ploidy	2.96
Buffer	RNase-free H <sub>2</sub> O
Storage Conditions	-80°C
Date of Manufacture	2023/04/27
Expiry	12 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*2 vials	1.00 µg/vial*2 vials	√
Concentration	60.00 ng/uL±10%	58.00 ng/uL by Nanodrop detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.8~2.2	1.91 by Nanodrop detection	√
RQN value	>9	9.82 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing	√
Copy number	Positive	763.21 copies/ng with bio-rad on-step Kit detection	

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*