


CERTIFICATE OF ANALYSIS

Product information

Name	ABL1(E1)-BCR(E15) Fusion	
Catalog No.	CBP20037R	
Lot No.	CBFA23091225X	
Format	RNA	
Left Gene	ABL1(E1)	BCR(E14)
Left Breakpoint	chr9:133589842:+	chr22:23632600:+
Right Gene	BCR(E15)	ABL1(E2)
Right Breakpoint	chr22:23634728:+	chr9:133729451:+
Ploidy	2.55	
Buffer	RNase-free H ₂ O	
Storage Conditions	-80°C	
Date of Manufacture	2023/09/12	
Expiry	12 months from the date of manufacture	

Quality Control

Test Items	Standard	Results	Passed
Quantity	0.30 µg/vial*1 vial	0.30 µg/vial*1 vial	√
Concentration	30.00 ng/uL±10%	33.00 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	2.02 by Nanodrop detection	√
RQN value	>8	9.38 by Qsep1 detection	
Genotype	Positive	 By Sanger Sequencing	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****