


CERTIFICATE OF ANALYSIS

BCR(E13)-ABL1(E2) Translocation

Product information

Name	BCR(E13)-ABL1(E2) Translocation
Catalog No.	CBP20061D
Lot No.	CGD04725188
Format	Genomic DNA
Left Gene	BCR(E13)
Left Breakpoint	N/A
Left Gene Copy Number	4
Right Gene	ABL1(E2)
Right Breakpoint	N/A
Right Gene Copy Number	4
Ploidy	2.53
%AF	N/A
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jan.-05-2022
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.82	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 5000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: BCR(E13)-ABL1(E2) Translocation </p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****