



CERTIFICATE OF ANALYSIS KMT2A-MLLT3 Fusion

Product information

Name	KMT2A-MLLT3 Fusion
Catalog No.	CBP20128R
Lot No.	CGD2021071004
Format	RNA
Left Gene	KMT2A
Left Breakpoint	chr11:118353210:+
Right Gene	MLLT3
Right Breakpoint	chr9:20365742:-
Ploidy	2.13
Buffer	RNase-free H2O
Storage Conditions	-80°C
Date of Manufacture	Jul.-10-2021
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	60 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.1	1.91	√
RNA electrophoresis	No Degradation	 M 1 M: Marker 1: KMT2A-MLLT3 Fusion	√
Genotype	Sanger Sequencing		√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****