

CERTIFICATE OF ANALYSIS

General information

Name	AI-Edigene® ETV6(E5)-NTRK3(E15) Fusion	
Catalog No.	CBP20140R	
Lot No.	CBFA22121222X	
Format	RNA	
Left Gene	ETV6(E5)	
Left Breakpoint	chr12:12022903:+	
Right Gene	NTRK3(E15)	
Right Breakpoint	chr15:88483984:-	
Ploidy	1.98	
Buffer	RNase-free H ₂ O	
Storage Conditions	-80°C	
Date of Manufacture	Dec12-2022	
Expiry	12 months from the date of manufacture	

Quality Control

Quality Control				
Test Items	Standard	Results	Passed	
Quantity	1.00 µg/vial*1 vial	1.00 μg/vial*1 vial		
Concentration	60.00 ng/uL±10%	58.60 ng/uL by Nanodrop detection		
OD_{260}/OD_{280}	1.8~2.2	1.99 by Nanodrop detection		
RQN value	>9	9.83 by Qsep1 detection	V	
Genotype	Positive	TOGGA GA AT A GCA GA TOTO CA GCA CA TTA A GA TO GA GA AT A GCA GA TOTO CA GCA CA TTA A GA TO GA GA GA TOTO CA GCA CA TTA A GA TOTO CA GCA CA CA TTA A GA TOTO C	√	
Copy Number	Positive	5032.17 copies/ng by DdPCR detection		

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:***

Date:****

Date:****

Tel: 4008-750-250

QQ: 4008-750-250

Email: sales@cobioer.com