

CERTIFICATE OF ANALYSIS

i i ouuce mitor mation			
Name	AI-Edigene® QKI(E6)-NTRK2(E14) Fusion		
Catalog No.	CBP20146R		
Lot No.	CBFA23060801x		
Format	RNA		
Left Gene	QKI (E6)		
Left Breakpoint	chr6:163984751:+		
Right Gene	NTRK2(E14)		
Right Breakpoint	chr9:87482158:+		
Ploidy	1.98		
Buffer	RNase-free H ₂ O		
Storage Conditions -80°C			
Date of Manufacture 2023/06/08			
Expiry 12 months from the date of manufacture			

Product information

Quality Control

Test Items	Standard	Results	Passed	
Quantity	0.50 µg/vial*1 vial	0.50 μg/vial*1 vial		
Concentration	60.00 ng/uL±10%	58.30 ng/uL by Nanodrop detection		
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.83 by Nanodrop detection		
RQN value	>9	9.94 by Qsep1 detection		
Genotype	Positive	CCTAGIGGIGIATIAGCCCCAGCCTCCGTTA CCTAGIGGIGIGIATIAGCCCCAGCCTCCGTTA MAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAA	\checkmark	
Copy number	Positive	471.06 copies/ng by DdPCR detection		

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 月時行 Date:**** 項望命 Date:****

Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com