


CERTIFICATE OF ANALYSIS

Product information

Name	MET Exon14 Skipping(E13-E15)
Catalog No.	CBP20171R
Lot No.	CBFA2023020513
Format	RNA
Left Gene	MET-E13
Left Breakpoint	chr7:116411708:+
Right Gene	MET-E15
Right Breakpoint	chr7:116414935:+
Ploidy	2.06
Buffer	RNase-free H ₂ O
Storage Conditions	-80°C
Date of Manufacture	2023/02/05
Expiry	12 months from the date of manufacture


Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	60.00 ng/uL±10%	57.40 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.87 by Nanodrop detection	√
RQN value	>9	9.17 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing	√
Copy Number	Positive	20176.18 copies/ng by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****