



# CERTIFICATE OF ANALYSIS

## MET Exon14 Skipping

### Product information

Name	MET Exon14 Skipping
Catalog No.	CBP20171R
Lot No.	CGD2021022501
Format	RNA
Left Gene	MET-E13
Left Breakpoint	chr7:116411708:+
Right Gene	MET-E15
Right Breakpoint	chr7:116414935:+
Ploidy	2.06
Buffer	RNase-free H2O
Storage Conditions	-80°C
Date of Manufacture	Feb.-25-2021
Expiry	12 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	60 ng/uL±10%	√
OD <sub>260</sub> /OD <sub>280</sub>	1.8~2.1	1.87	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: MET-MET Fusion</p>	√
Genotype	Sanger Sequencing		√
Copy Number	DdPCR	27279.64copies/ng	√

#### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*