


CERTIFICATE OF ANALYSIS

Product information

Name	TPM3(E7)-NTRK1(E10) Fusion
Catalog No.	CBP20178R
Lot No.	CBFA22120605X
Format	RNA
Left Gene	TPM3(E7) NM_001043351.2
Left Breakpoint	chr1:154142876:-
Right Gene	NTRK1 NM_002529.3
Right Breakpoint	chr1:156844363:+
Ploidy	1.93
Buffer	RNase-free H ₂ O
Storage Conditions	-80°C
Date of Manufacture	Dec.-06-2022
Expiry	12 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*3 vials	1.00 µg/vial*3 vials	√
Concentration	60.00 ng/uL±10%	61.20 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.84 by Nanodrop detection	√
RQN value	>9	9.8 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing	√
Copy number	Positive	5235.86 copies/ng by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****