

CERTIFICATE OF ANALYSIS

Product information

Name	AI-Edigene® EML4(E6)-ALK(E20) V3a Fusion			
Catalog No.	CBP20184R			
Lot No.	CBFA22112510X			
Format	RNA			
Left Gene	EML4(E6)			
Left Breakpoint	chr2:42491871:+			
Right Gene	ALK(E20)			
Right Breakpoint	chr2:29446394:-			
Ploidy	1.98			
Buffer	RNase-free H ₂ O			
Storage Conditions	-80°C			
Date of Manufacture	Nov25-2022			
Expiry	12 months from the date of manufacture			

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	
Concentration	$60.00 \text{ ng/uL} \pm 10\%$	60.00 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.84 by Nanodrop detection	V
RQN value	>9	9.38 by Qsep1 detection	
Genotype	Positive	By Sanger Sequencing	V
Copy Number	Positive	454.36 copies/ng by DdPCR detection	V

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:***

Date:****

Date:****

Tel: 4008-750-250

QQ: 4008-750-250 Email: sales@cobioer.com