

CERTIFICATE OF ANALYSIS

Product information

Name	AI-Edigene® EML4(E6)-ALK(E20) V3b Fusion
Catalog No.	CBP20185R
Lot No.	CGFA23081902X
Format	RNA
Left Gene	EML4(E6)
Left Breakpoint	chr2:42492091:+
Right Gene	ALK(E20)
Right Breakpoint	chr2:29446394:-
Ploidy	1.98
Buffer	RNase-free H ₂ O
Storage Conditions	-80°C
Date of Manufacture	2023/08/19
Expiry	12 months from the date of manufacture

Quality Control

Quanty Control			
Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	
Concentration	$60.00 \text{ ng/uL} \pm 10\%$	59.80 ng/uL by Nanodrop detection	V
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	2.03 by Nanodrop detection	V
RQN value	>9	9.78 by Qsep1 detection	V
Genotype	Positive		√
		By Sanger Sequencing	
Copy number	Positive	442.09 copies/ng by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:***

Date:***

Date:***

Tel: 4008-750-250

QQ: 4008-750-250

Email: sales@cobioer.com