


CERTIFICATE OF ANALYSIS

Product information

Name	KMT2A-MLLT3(E10-E6,E9-E6) Fusion	
Catalog No.	CBP20191R	
Lot No.	CBFA23091230X	
Format	RNA	
Left Gene	KMT2A	KMT2A
Left Breakpoint	chr11:118355690:+	chr11:118355029:+
Right Gene	MLLT3	MLLT3
Right Breakpoint	chr9:20365742:-	chr9:20365742:-
Ploidy	2.17	
Buffer	RNase-free H ₂ O	
Storage Conditions	-80°C	
Date of Manufacture	2023/09/12	
Expiry	12 months from the date of manufacture	



Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	60.00 ng/uL±10%	66.00 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.96 by Nanodrop detection	√
RQN value	>8	9.25 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing for KMT2A E9-MLLT3 E6	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****