



CERTIFICATE OF ANALYSIS

KMT2A-PHLDB1 Translocation

Product information

Name	KMT2A-PHLDB1 Translocation
Catalog No.	CBP20201D
Lot No.	CGD05741372
Format	Genomic DNA
Left Gene	KMT2A
Left Breakpoint	chr11:118365055:-
Left Gene Copy Number	5
Right Gene	PHLDB1
Right Breakpoint	chr11:118499371:+
Right Gene Copy Number	5
Ploidy	3.16
%AF	Unknown
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jun.-14-2021
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.82	√
DNA electrophoresis	No Degradation	 <p style="text-align: center;">M 1</p> <p>M: Marker 1: KMT2A-PHLDB1 Translocation</p>	√
Genotype	Positive		√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****