


CERTIFICATE OF ANALYSIS FGFR2 Amplification Reference Standard Product information

Name	FGFR2 Amplification Reference Standard
Catalog No.	CBP40003
Lot No.	CGD97687908
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	6 (DdPCR)
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Dec.-12-2019
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.78	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ... ng/5μl</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>5000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: FGFR2 Amplification Reference Standard</p>	√
Copy number	DdPCR	CN=6.31	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****