


CERTIFICATE OF ANALYSIS

MDM2 Amplification Reference Standard

Product information

Name	MDM2 Amplification Reference Standard
Catalog No.	CBP40011
Lot No.	CGD05399410
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	14 (NGS)
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Aug.-14-2020
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	35 ng/uL	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.79	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng5ul 10000 40 8000 40 6000 40 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: MDM2 Amplification Reference Standard </p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****