


CERTIFICATE OF ANALYSIS

Product information

Name	NF2 Loss Reference Standard
Catalog No.	CBP40027
Lot No.	CBFA22120707
Format	Genomic DNA
Copy number variation	Loss
Measured CNV	0 (NGS)
Ploidy	3.32
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2022/12/07
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	39.20 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.84 by Nanodrop detection	√
DNA electrophoresis	No Degradation	1kb ladder  M: Marker 1: NF2 Loss Reference Standard	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****