


CERTIFICATE OF ANALYSIS

FGFR1 Amplification Reference Standard

Product information

Name	FGFR1 Amplification Reference Standard
Catalog No.	CBP40037
Lot No.	CGD54143353
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	20 (DdPCR)
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Aug.-27-2019
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.79	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: FGFR1 Amplification Reference Standard</p>	√
Copy Number	DdPCR	CN=20.31	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****