


CERTIFICATE OF ANALYSIS

MET Amplification Reference Standard

Product information

Name	MET Amplification Reference Standard
Catalog No.	CBP40043
Lot No.	GBD80175405
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	10 (DdPCR)
Ploidy	2.06
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jul.-15-2020
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.80	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: MET Amplification Reference Standard</p>	√
Copy Number	DdPCR	CN=10.6	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****