CERTIFICATE OF ANALYSIS

MET Amplification Reference Standard

Product information

Name	MET Amplification Reference Standard
Catalog No.	CBP40044
Lot No.	CGD49691621
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	38.4 (DdPCR)
Buffer	Tris-EDTA
Storage Conditions	-20℃
Date of Manufacture	Apr.-26-2019
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD_{260}/OD_{280}	1.7~2.0	1.80	√
DNA electrophoresis	No Degradation	 M: Marker 1: MET Amplification Reference Standard	√
Copy Number	DdPCR	CN=37	√

Caution:
For research use only.
If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:****

Date:****

No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu
www.cobioer.com
Tel: 4008-750-250
QQ: 4008-750-250
Email: sales@cobioer.com