


CERTIFICATE OF ANALYSIS

Product information

Name	MYC Amplification Reference Standard
Catalog No.	CBP40045
Lot No.	CBFA22112220
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	59 (DdPCR)
Ploidy	3.37
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2022/11/22
Expiry	36 months from the date of manufacture



Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	39.30 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.82 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: MYC Amplification Reference Standard</p>	√
Copy Number	Amplification	CN=58.56 by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****