


## CERTIFICATE OF ANALYSIS

### Product information

Name	CDKN2B Loss Reference Standard
Catalog No.	CBP40055
Lot No.	CBFA22101902
Format	Genomic DNA
Copy number variation	Loss
Measured CNV	0 (NGS)
Ploidy	3.46
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2022/10/19
Expiry	36 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*6 vials	1.00 µg/vial*6 vials	√
Concentration	40.00 ng/uL±10%	38.70 ng/uL by Qubit 3.0 detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.73 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>           1kb ladder            bp ng/5µl            10000 40            8000 40            6000 40            4000 40            3000 40            2000 100            1500 40            1000 40            500 40            1% agarose            M            1            M: Marker            1: CDKN2B Loss Reference Standard         </p>	√

**Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*