


CERTIFICATE OF ANALYSIS

RAD51B Loss Reference Standard

Product information

Name	RAD51B Loss Reference Standard
Catalog No.	CBP40079
Lot No.	CGD04812319
Format	Genomic DNA
Copy number variation	Loss
Measured CNV	0 (DdPCR)
Ploidy	2.97
Buffer	Water
Storage Conditions	-20°C
Date of Manufacture	Apr.-12-2021
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	10ug/vial*3vials	10ug/vial*3vials	√
Concentration	Qubit 3.0	20 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.95	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: RAD51B Loss Reference Standard</p>	√
Copy number	DdPCR	CN=0	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****