


CERTIFICATE OF ANALYSIS

Product information

Name	MSH2 Loss Reference Standard
Catalog No.	CBP40087
Lot No.	GBD87074742
Format	Genomic DNA
Copy number variation	Loss
Measured CNV	1 (NGS)
Ploidy	2.04
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Nov.-16-2021
Expiry	36 months from the date of manufacture



Quality Control

Test Items	Standard	Results	Passed
Quantity	5.00 µg/vial*1 vial	5.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	38.40 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.83 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1:MSH2 Loss Reference Standard</p>	√
Copy number	Loss	CN=1.06 by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****