


CERTIFICATE OF ANALYSIS

EGFR Amplification Reference Standard

Product information

Name	EGFR Amplification Reference Standard
Catalog No.	CBP40093
Lot No.	CGD57913696
Format	Genomic DNA
Copy number variation	Amplification
Measured CNV	11.49(DdPCR)
Ploidy	3.3
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Dec.-04-2019
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.86	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: EGFR Amplification Reference Standard</p>	√
Copy Number	Amplification	CN=10.6 by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****