


CERTIFICATE OF ANALYSIS

General information

Name	gDNA of HCT116
Catalog No.	CBP60028
Lot No.	CGD20114371
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jun.-02-2020
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	5ug/vial*1vial	5ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD260/OD280	1.75~1.9	1.80	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 5000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: gDNA of HCT116 </p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****