


CERTIFICATE OF ANALYSIS

General information

Name	tTMB-P1(5.37) 配对标准品
Catalog No.	CBP80001-1
Lot No.	CGFA23050607
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/05/06
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*4 vials	1.00 µg/vial*4 vials	√
Concentration	40.00 ng/uL±10%	42.40 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.80 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5µl 10000 40 8000 40 6000 40 5000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M M: Marker 1: tTMB-P1(5.37) 配对标准品 </p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date: ****

 Date: ****