CERTIFICATE OF ANALYSIS

General information

Name	tTMB-P4(21.09) Reference Standard_Normal		
Catalog No.	CBP80001-4		
Lot No.	CBFA22010512		
Format	Genomic DNA		
Buffer	Tris-EDTA		
Storage Conditions	-20°C		
Date of Manufacture	Jan05-2022		
Expiry	36 months from the date of manufacture		

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	
Concentration	40.00 ng/uL±10%	39.00 ng/uL by Qubit 3.0 detection	
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.93 by Nanodrop detection	
DNA electrophoresis	No Degradation	Ikb ladder Ikb ladder Ikb ladder M 1 M: Marker 1: tTMB-P4(21.09) Reference Standard_Normal	\checkmark

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date: **** 項項子 Date: ****

No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com