


# CERTIFICATE OF ANALYSIS

## General information

Name	tTMB-P4(21.09) 配对标准品
Catalog No.	CBP80001-4
Lot No.	CBFA22120704
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Dec.-07-2022
Expiry	36 months from the date of manufacture

## Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*5 vials	1.00 µg/vial*5 vials	√
Concentration	40.00 ng/uL±10%	39.00 ng/uL by Qubit 3.0 detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.93 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>           1kb ladder            bp ng/5µl            10000 40            8000 40            6000 40            5000 100            4000 40            3000 40            2000 100            1500 40            1000 40            500 40            1% agarose            M            M: Marker            1: tTMB-P4(21.09) 配对标准品         </p>	√

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date: \*\*\*\*

 Date: \*\*\*\*