


# CERTIFICATE OF ANALYSIS

## General information

Name	tTMB-P5(27.15) Reference Standard_Normal
Catalog No.	CBP80001-5
Lot No.	GBD36855770
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Sep.-06-2020
Expiry	36 months from the date of manufacture

## Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD260/OD280	1.75~1.9	1.81	√
DNA electrophoresis	No Degradation	 <p>           1kb ladder            bp ng/5ul            10000 40            8000 40            6000 40            5000 100            4000 40            3000 40            2000 100            1500 40            1000 40            500 40            1% agarose            M 1            M: Marker            1: tTMB-P5(27.15) Reference Standard Normal         </p>	√

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*