CERTIFICATE OF ANALYSIS

General information

Name	MSI-H-U6配对标准品
Catalog No.	CBP80002-10
Lot No.	CGFA23050603
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/05/06
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results			Passed
Quantity	0.50 μg/vial*1 vial	0.50 μg/vial			
Concentration	40.00 ng/uL±10%	38.80 ng/uL			
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.84 by Nan			
DNA electrophoresis	No Degradation	1kb ladder 1kb ladder 1kb ladder 1kb ladder 1kb ladder 1000 40 1000 4	\checkmark		
MSI/MSS status		MSS by PCR detection			
		M19-11060			
	MSS	Marker	Size 1	Size 2	
		NR21	131.9		
		Bat26	176.32		\checkmark
		Bat25	118.87		
		NR27	155.54		
		NR24	131.99		
		Mono27	171.78		

Caution:

For research use only. If you have any questions about the Certificate of Analysis, please contact us.

No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com





No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com