CERTIFICATE OF ANALYSIS

General information

Name	MSI-H-U8配对标准品
Catalog No.	CBP80002-12
Lot No.	CBFA23020306
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/02/03
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	0.50 µg/vial*1 vial	0.50 μg/vial*1 vial	
Concentration	40.00 ng/uL±10%	41.30 ng/uL by Qubit 3.0 detection	
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.87 by Nanodrop detection	
DNA electrophoresis	No Degradation	Ikb ladder	\checkmark
MSI/MSS status		MSS by PCR detection	
		M19-11060	
	MSS	Marker Size 1 Size 2	1
		NR21 131.9	
		Bat26 176.32] √
		Bat25 118.87	
		NR27 155.54	
		NR24 131.99	
		Mono27 171.78	

Caution:

For research use only. If you have any questions about the Certificate of Analysis, please contact us.

No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com





No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com