

CERTIFICATE OF ANALYSIS

General information

Name	MSS-P3 标准品
Catalog No.	CBP80002-3
Lot No.	CBFA22081101
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2022/08/11
Expiry	36 months from the date of manufacture

Quality Control

Quality Control						
Test Items	Standard	Results			Passed	
Quantity	0.50 μg/vial*1 vial	0.50 μg/vial	V			
Concentration	40.00 ng/uL±10%	41.10 ng/uL	V			
OD_{260}/OD_{280}	1.7~2.1	1.86 by Nan	√			
DNA electrophoresis	No Degradation	1000 40 1000 40 100	M 1 M: Marker 1: MSS-P3 标准品			
MSI/MSS status	MSS	MSS by PCI	MSS by PCR detection			
		M19-9745				
		Marker	Size 1	Size 2		
		NR21	131.6			
		Bat26	175.26			
		Bat25	118.65			
		NR27	155.3			
		NR24	131.89			
		Mono27	171.7			

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date: ****

1**季**1日和 Date: ****

Tel: 4008-750-250

QQ: 4008-750-250

Email: sales@cobioer.com