


CERTIFICATE OF ANALYSIS

General information

Name	MSI-H-U1(100%)标准品
Catalog No.	CBP80002-5
Lot No.	CBFA22123004
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Dec.-30-2022
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed																								
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√																								
Concentration	40.00 ng/uL±10%	39.00 ng/uL by Qubit 3.0 detection	√																								
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.86 by Nanodrop detection	√																								
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5µl 10000 40 8000 40 6000 40 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: MSI-H-U1(100%)标准品 </p>	√																								
MSI/MSS status	MSI	MSI by PCR detection <table border="1"> <thead> <tr> <th colspan="3">M19-11062</th> </tr> <tr> <th>Marker</th> <th>Size 1</th> <th>Size 2</th> </tr> </thead> <tbody> <tr> <td>NR21</td> <td>120.92</td> <td></td> </tr> <tr> <td>Bat26</td> <td>166.04</td> <td></td> </tr> <tr> <td>Bat25</td> <td>109.62</td> <td></td> </tr> <tr> <td>NR27</td> <td>143.22</td> <td></td> </tr> <tr> <td>NR24</td> <td>125.79</td> <td></td> </tr> <tr> <td>Mono27</td> <td>161.65</td> <td></td> </tr> </tbody> </table>	M19-11062			Marker	Size 1	Size 2	NR21	120.92		Bat26	166.04		Bat25	109.62		NR27	143.22		NR24	125.79		Mono27	161.65		√
M19-11062																											
Marker	Size 1	Size 2																									
NR21	120.92																										
Bat26	166.04																										
Bat25	109.62																										
NR27	143.22																										
NR24	125.79																										
Mono27	161.65																										
MSI/MSS	MSI	100% by NGS-Panel	√																								

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date: 03/17/2023

 Date: 03/17/2023