CERTIFICATE OF ANALYSIS

General information

Name	MSI-H-U2(98.08%) 配对标准品
Catalog No.	CBP80002-6
Lot No.	CBFA22101201
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Oct12-2022
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results			Passed	
Quantity	1.00 μg/vial*1 vial	1.00 µg/vial				
Concentration	40.00 ng/uL±10%	38.70 ng/uL				
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.84 by Nan	√			
DNA electrophoresis	No Degradation	1kb ladder 1kb ladder 1kb ladder 1kb ladder 1kb ladder 1000 40 1000 400 1000 40 1000 400 1000 40 1000 40 1000 40 1000	\checkmark			
MSI/MSS status	MSS	MSS by PCR detection				
		M19-11060				
		Marker	Size 1	Size 2		
		NR21	131.9			
		Bat26	176.32			
		Bat25	118.87			
		NR27	155.54			
		NR24	131.99			
		Mono27	171.78			

Caution:


For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com



Your Intimate Partner in Biosciences!



No.222 Changhong Road, Yuhua District,Nanjing,Jiangsu www.cobioer.com Tel: 4008-750-250 QQ: 4008-750-250 Email: sales@cobioer.com