Tel: 4008-750-250

QQ: 4008-750-250 Email: sales@cobioer.com



CERTIFICATE OF ANALYSIS

General information

Name	MSI-H-U3标准品
Catalog No.	CBP80002-7
Lot No.	CGFA23051003
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/05/10
Expiry	36 months from the date of manufacture

Quality Control

Quality Control							
Test Items	Standard	Results	Passed				
Quantity	1.00 μg/vial*1 vial	1.00 µg/vial					
Concentration	40.00 ng/uL±10%	43.80 ng/uL					
OD_{260}/OD_{280}	1.7~2.1	1.93 by Nan	√				
DNA electrophoresis	No Degradation	Ikb ladder 10000 40 10000 4			V		
MSI/MSS status	MSI	MSI-H by PCR detection					
		M19-11064					
		Marker	Size 1	Size 2			
		NR21	123.53				
		Bat26			√		
		Bat25	110.45	114.22			
		NR27	142.98				
		NR24	125.79	127.59			
		Mono27	162.21				

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.



Your Intimate Partner in Biosciences!

Certified by: Date: ****

Date: ****

Date: ****

Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com