Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com



CERTIFICATE OF ANALYSIS

General information

Name	MSI-H-U4配对标准品
Catalog No.	CBP80002-8
Lot No.	CBFA23121504
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/12/15
Expiry	36 months from the date of manufacture

Quality Control

Quality Control							
Test Items	Standard	Results			Passed		
Quantity	1.00 μg/vial*20 vials	1.00 μg/vial*20 vials			√		
Concentration	40.00 ng/uL±10%	44.00 ng/uL by Qubit 3.0 detection					
OD_{260}/OD_{280}	1.7~2.1	1.77 by Nanodrop detection					
DNA electrophoresis	No Degradation	1kb ladder			√		
MSI/MSS status	MSS	MSS by PCR detection					
		M19-11060					
		Marker	Size 1	Size 2			
		NR21	131.9		√		
		Bat26	176.32				
		Bat25	118.87				
		NR27	155.54				
		NR24	131.99				
		Mono27	171.78				

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.





Certified by: Date:***

Date:****

Date:****

Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com