



CERTIFICATE OF ANALYSIS

Human Methylation-Normal Reference Standard(Normal sample)

General information

Name:	Human Methylation-Normal Reference Standard(Normal sample)
Catalog No.	CBP90035-4
Lot No.	GBD47739225
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Aug.-31-2021
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.88	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: Human Methylation-Normal Reference Standard(Normal sample)</p>	√
Methylation	Normal		√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****