



CERTIFICATE OF ANALYSIS

CYP2C9*2 Reference Standard

Product information

Name	CYP2C9*2 Reference Standard
Catalog No.	CBPA0011
Lot No.	GBD80555547
Format	Genomic DNA
DNA Change	NM_000771.3_c.430C>T
AA Change	p.R144C
Chr position (GRCh38)	NC_000010.11_g.94942290C>T
Zygosity	Heterozygous
Allelic Frequency	50%
Copy Number	2
Ploidy	2
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	May-19-2020
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.85	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>I</p> <p>M: Marker</p> <p>1: CYP2C9*2 Reference Standard</p>	√
Genotype	Sanger Sequencing		√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****