



CERTIFICATE OF ANALYSIS

Product information

Name	CYP2C9*2 Reference Standard
Catalog No.	CBPA0011
Lot No.	GBD88870952
Format	Genomic DNA
DNA Change	NM_000771.3 c.430C>T
AA Change	p.R144C
Chr position (GRCh38)	NC_000010.11 g.94942290C>T
Zygoty	Heterozygous
%AF	50.00%
Copy Number	2
Ploidy	2
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Jan.-27-2022
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	40.90 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.79 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: CYP2C9*2 Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****