



CERTIFICATE OF ANALYSIS

Product information

Name	MTHFR 677C>T Reference Standard
Catalog No.	CBPA0019
Lot No.	CBFA23090503
Format	Genomic DNA
DNA Change	NM_005957.5_c.667C>T
AA Change	p.A222V
Chr position (GRCh38)	NC_000001.11_g.11796321G>A
Zygosity	Heterozygous
%AF	50.00%
Copy Number	2
Ploidy	2
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/09/05
Expiry	36 months from the date of manufacture


Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	42.60 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.84 by Nanodrop detection	√
DNA electrophoresis	No Degradation	<p>1kb ladder</p>  <p>M: Marker 1: MTHFR 677C>T Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****