



CERTIFICATE OF ANALYSIS

Product information

Name	SLCO1B1*15 Reference Standard
Catalog No.	CBPA0057
Lot No.	GBD64368020
Format	Genomic DNA
DNA Change	NM_006446.4:c.521T>C
AA Change	p.V174A
Chr position (GRCh38)	NC_000012.12:g.21178615T>C
Zygoty	Homozygous
%AF	100.00%
Copy Number	2
Ploidy	2
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2021/05/11
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*5 vials	1.00 µg/vial*5 vials	√
Concentration	40.00 ng/uL±10%	41.50 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.93 by Nanodrop detection	√
DNA electrophoresis	No Degradation	1kb ladder  M: Marker I: SLCO1B1*15 Reference Standard	√
Genotype	Positive	 By sanger sequencing	√


Caution:

No.222 Changhong Road, Yuhua
 District, Nanjing, Jiangsu
 www.cobioer.com

Tel: 4008-750-250
 QQ: 4008-750-250
 Email: sales@cobioer.com

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****