Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com



## **CERTIFICATE OF ANALYSIS**

## **Product information**

3.7	GTD2 G12VIG to D G 1 1
Name	GJB2 p.G12Vfs*2 Reference Standard
Catalog No.	CBPD0021
Lot No.	CBFA23081406
Format	Genomic DNA
DNA Change	c.35delG
AA Change	p.G12Vfs*2
Zygosity	Heterozygous
%AF	50.00%
Transcript	NM_004004.6
Chr position (GRCh37)	chr13-20763686-C-
Copy Number	2
Ploidy	2
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/08/14
Expiry	36 months from the date of manufacture

**Quality Control** 

Quanty Control			
Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	
Concentration	40.00 ng/uL±10%	40.00 ng/uL by Qubit 3.0 detection	
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.95 by Nanodrop detection	V
DNA electrophoresis	No Degradation	1kb ladder	√
Genotype	Positive	By sanger sequencing	<b>√</b>

## **Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.





Certified by: Date:\*\*\*

Date:\*\*\*\*

Date:\*\*\*\*

Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com