

CERTIFICATE OF ANALYSIS

2019-nCov-N Pseudovirus Standard Reference

Product information

Name	2019-nCov-N Pseudovirus Standard Reference
Catalog No.	CBV30002
Lot No.	CGD2022032201
Format	Virus
Gene	Gene N
Sequence	1260bp (28274-29533): ATGTCTGATAATGGACCCCCAAAATCAGCGAAATGCACCCCGCATTACGTTTGGTGGACCC TCAGATTCAACTGGCAGTAACCAGAATGGAGAACGCAGTGGGGCGCGATCAAAACAACGT CGGCCCAAGGTTTACCCAATAATACTGCGTCTTGGTTACCGCTCTCACTCAACATGGC AAGGAAGACCTTAAATTCCTCGAGGACAAGGCGTTCCAATTAACACCAATAGCAGTCCA GATGACCAAAATGGCTACTACCGAAGAGCTACCAGACGAATTCGTGGTGGTGACGGTAAA ATGAAAGATCTCAGTCCAAGATGGTATTTCTACTACCTAGGAAGTGGGCCAGAAGCTGGA CTTCCCTATGGTGCTAACAAGACGGCATCATATGGGTTGCAACTGAGGGAGCCTTGAAT ACACCAAAAGATCACATTGGCACCCTGCAATCCTGCTAACAATGCTGCAATCGTGCTACAA CTTCCTCAAGGAACAACATTGCCAAAAGGCTTCTACGCAGAAGGGAGCAGAGGGCGGCAGT CAAGCCTCTTCTCGTTCCTCATCACGTAGTCGCAACAGTTCGAAGAAATCAACTCCAGGC AGCAGTAGGGGAACCTTCTCTGCTAGAATGGCTGGCAATGGCGGTGATGCTGCTCTTGCT TTGCTGCTGCTTGACAGATTGAACCAGCTTGAGAGCAAAATGTCTGGTAAAGGCCAACAA CAACAAGGCCAAACTGTCAC TAAGAAATCTGCTGCTGAGGCTTCTAAGAAGCCTCGGC AA AAACGTACTGCCACTAAAGCATAACAATGTACACAAGCTTTCGGCAGACGTTGGTCCAGAA CAAACCCAAGGAAATTTGGGGACCAGGAACTAATCAGACAAGGAACTGATTACAAACAT TGGCCGCAAATGCACAATTTGCCCCAGCGCTTACGCGTCTTCGGAATGTCGCGCATT GGCATGGAAGTCACACCTTCGGGAACGTGGTTGACCTACACAGGTGCCATCAAATGGAT GACAAAGATCCAAATTTCAAAGATCAAGTCATTTTGCTGAATAAGCATATTGACGCATAC AAAACATTTCCACCAACAGAGCCTAAAAAGGACAAAAAGAAGAAGGCTGATGAAACTCAA GCCTTACCGCAGAGACAGAAGAAACAGCAAACTGTGACTCTTCTTCCCTGCTGCAGATTTG GATGATTTCTCCAAACAATTCGAACAATCCATGAGCAGTGCTGACTCAACTCAGGCCTAA
NCBI No.	NC_045512.2
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	03.22.2022
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*2vials	1ml/vial*2vials	√
Copy Number	>1x10e7copies/ml	2.48*10e7copies/ml	√
DNA Residue	<2%	0.90%	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****