

## CERTIFICATE OF ANALYSIS

### 2019-nCov-E Pseudovirus Standard Reference

#### Product information

Name	2019-nCov-E Pseudovirus Standard Reference
Catalog No.	CBV30003
Lot No.	CGD2020032301
Format	Virus
Gene	Gene E
Sequence	228bp ATGTACTCATTTCGTTTTCGGAAGAGACAGGTACGTTAATAGTTAATAGCGTACTTCTTTTT CTTGCTTTTCGTGGTATTCTTGCTAGTTACACTAGCCATCCTTACTGCGCTTCGATTGTGT GCGTACTGCTGCAATATTGTTAACGTGAGTCTTGTA AACCTTCTTTTTTACGTTTACTCT CGTGTTAAAAATCTGAATTCCTTAGAGTTCTGATCTTCTGGTCTAA
NCBI No.	NC_045512.2
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	03.23.2020
Expiry	12 months from the date of manufacture

#### Quality Control

Test Items	Standard	Results	Passed
Quantity	100ul/vial*1vial	100ul/vial*1vial	√
Copy Number	>1x10 <sup>7</sup> copies/ml	4.0x10 <sup>8</sup> copies/ml	√

#### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*