

# CERTIFICATE OF ANALYSIS

## SARS-N Pseudovirus Standard Reference

### Product information

Name	SARS-N Pseudovirus Standard Reference
Catalog No.	CBV30006
Lot No.	CGD2020092511
Format	Virus
Gene	Gene N
Sequence	<p>1207bp:</p> <p>ATGTCGTGATAATGGACCCCAATCAAACCAACGCTAGTGCCCCCGCATTACATTTGGTGGACCCACAGATTCAACTGACAATAACCAGAATGGAGGACGCAATGGGGCAAGGCCAAAACAGCGCCGACCCCAAGGTTTACCCAATAATACTGCGTCTTGGTTCACAGCTCTCACTCAGCATGGCAAGGAGGAACTTAGATTCCCTCGAGGCCAGGGCGTTCCAATCAACACCAATAGTGGTCCAGATGACCAAATTGGCTACTACCGAAGAGCTACCCGACGAGTTCTGGTGGTGACGGCAAAATGAAAGAGCTCAGCCCCAGATGGTACTTCTATTACCTAGGAACTGGCCAGAAGCTTCACTTCCCTACGGCGCTAACAAAGAAGGCATCGTATGGGTGCAACTGAGGGAGCCTTGAAATACCCCAAAGACCACATTGGCACCCGCAATCCTAATAACAATGCTGCCACCGTGCTCAACTTCTCAAGGAACAACATTGCCAAAAGGCTTCTACGCAGAGGGAAGCAGAGGGCGGAGTCAAGCCTTCTCGCTCCTCATCACGTAGTCGCGGTAATTCAAGAAATCAACTCCTGGCAGCAGTAGGGGAAATTTCTCCTGCTCGAATGGCTAGCGGAGGTGGTAAAAGTGGCCTCGCCTATTGCTGCTAGACAGATTGAACCAGCTTGAGAGCAAAGTTTCTGGTAAAGGCCAACACAACAAGGCCAACTGTCACTAAGAAATCTGCTGCTGAGGCATCTAAAAGCCTCGCCAAAACGCTACTGCCACAAAACAGTACAACGTCCTCAAGCATTTGGGAGACGTGGTCCGAACAAAACCAAGGAAATTTCCGGGGACCAAGACCTAATCAGACAAGGAACTGATTACAAACATTGGCCGCAATTTGCACAATTTGCTCCAAGTGCCTCTGCATTCTTTGGAATGTCACGATTTGGCATGGAAGTACACCTTCGGGAACATGGCTGACTTATCATGGAGCCATTAATTTGATGACAAAGATCCACAATTCAAAGACAACGTCATACTGCTGAACAAGCACATTGACGCAATACAAAACATTTCCACCAACAGAGCCTAAAAAGGACAAAAAGAAAAGACTGATGAAGCTCAGCCTTTGCCGAGAGACAAAAGAAGCAGCCCACTGTGACTCTTCTTCTGCGGCTGACATGGATGATTTCTCCAGACAACCTTCAAAATTCATGAGTGGAGCTTCTGCTGATTCAACTCAGGCATAA</p>
NCBI No.	NC_004718.3
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	09.25.2020
Expiry	12 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*1vial	1ml/vial*1vial	√
Copy Number	>1x10 <sup>7</sup> copies/ml	>1x10 <sup>7</sup> copies/ml	√

#### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*