

CERTIFICATE OF ANALYSIS

SARS-N Pseudovirus Standard Reference

Product information

Name	SARS-N Pseudovirus Standard Reference
Catalog No.	CBV30006
Lot No.	CGD2022022508
Format	Virus
Gene	Gene N
Sequence	<p>1207bp:</p> <p>ATGTC TGATAATGGACCCCAATCAAACCAACG TAGTGCCCCCGCATTACATTTGGTGGACCCACAGATTC AACTGACAATAACCAGAATGGAGGACGCAATGGGGCAAGGCCAAAACAGCGCCGACCCCAAGGTTTACCCAATAATACTGCGTCTTGGTTCACAGCTCTCACTCAGCATGGCAAGGAGGAACTTAGATTCCCTCGAGGCCAGGGCGTTCCAATCAACACCAATAGTGGTCCAGATGACCAAATTGGCTACTACCGAAGAGCTACCCGACGAGTTCTGTGGTGGTGACGGCAAAATGAAAGAGCTCAGCCCCAGATGGTACTTCTATTACCTAGGAACTGGCCAGAAGCTTCACTTCCCTACGGCGCTAACAAAGAAGGCATCGTATGGGTGCAACTGAGGGAGCCTTG AATACACCCAAAGACCACATTGGCACCCGCAATCCTAATAACAATGCTGCCACCGTGCTA CAACTTCTCAAGGAACAACATTGCCAAAAGGCTTCTACGCAGAGGGAAGCAGAGGGCGGC AGTCAAGCCTCTTCTCGCTCCTCATCACGTAGTCGCGGTAATTCAAGAAATCAACTCCT GGCAGCAGTAGGGGAAATTTCTCCTGCTCGAATGGCTAGCGGAGGTGGTAAAAGTGCCTC GCGCTATTGCTGCTAGACAGATTGAACCAGCTTGAGAGCAAAGTTTCTGGTAAAGGCCAA CAACAACAAGGCCAAACTGTCACTAAGAAATCTGCTGCTGAGGCATCTAAAAGCCTCGC CAAAACGTTACTGCCACAAAACAGTACAACGTCCTCAAGCATTGGGAGAGCTGGTCCA GAACAAACCCAAAGGAAATTTCCGGGGACCAAGACCTAATCAGACAAGGAACTGATTACAAA CATTGGCCGCAAATTCACAAATTTGCTCCAAGTGCCTCTGCATTCTTTGGAATGTCACGC ATTGGCATGGAAGTACACCTTCGGGAACATGGCTGACTTATCATGGAGCCATTAAATTTG ATGACAAAGATCCACAATTCAAAGACAACGTCATACTGCTGAACAAGCACATTGACGCA TACAAAACATTTCCACCAACAGAGCCTAAAAAGGACAAAAAGAAAAGACTGATGAAGCT CAGCCTTTGCCGAGAGACAAAAGAAGCAGCCCACTGTGACTCTTCTTCTCGCGGCTGAC ATGGATGATTTCTCCAGACAACCTTCAAAATTCATGAGTGGAGCTTCTGCTGATTCAACT CAGGCATAA</p>
NCBI No.	NC_004718.3
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	02.25.2022
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*1 vial	1ml/vial*1 vial	√
Copy Number	>1x10 ⁷ copies/ml	1.54*10 ⁷ copies/ml	√
DNA Residue	<1%	0.90%	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****