

CERTIFICATE OF ANALYSIS

MERS-N Pseudovirus Standard Reference

Product information

Name	MERS-N Pseudovirus Standard Reference
Catalog No.	CBV30009
Lot No.	CGD2021022503
Format	Virus
Gene	Gene N
Sequence	<p>1242bp:</p> <p>ATGGCATCCCCCTGCTGCACCTCGTGCTGTTTCCTTTGCCGATAACAATGATATAACAAAT ACAAACCTATCTCGAGGTAGAGGACGTAATCCAAAACCACGAGCTGCACCAAATAACACT GTCTCTTGGTACACTGGGCTTACCCAACACGGGAAAGTCCCTTACCCTTCCACCTGGG CAGGGTGTACCTCTTAATGCCAATTCTACCCTGCGCAAAATGCTGGGTATTGGCGGAGA CAGGACAGAAAATTAATACCGGAATGGAATTAAGCAACTGGCTCCCAGGTGGTACTTC TACTACACTGGAAGTGGACCCGAAGCAGCACTCCCATTCCGGGCTGTTAAGGATGGC ATC GTTTGGGTCCATGAAGATGGCGCCACTGATGCTCCTTCAACTTTTGGGACGCG GAACCCCT AACAATGATTCAGCTATTGTTACACAATTCGCGCCCGGTACTAAGCTTCT AAAAACTTC CACATTGAGGGGACTGGAGGCAATAGTCAATCATCTTCAAGAGCCTCT AGCTTAAGCAGA AACTCTTCCAGATCTAGTTACAAGTTCAAGATCAGGAACTCTACCC GCGGCACCTTCT CCAGGTCCATCTGGAATCGGAGCAGTAGGAGGTGATCTACTTTAC CTTGATCTTCTGAAC AGACTACAAGCCCTTGAGTCTGGCAAAGTAAAGCAATCGCAG CCAAAGTAATCACTAAG AAAGATGCTGCTGCTGCTAAAATAAGATGCGCCACAAGC GCACTTCCACCAAAGTTTC AACATGGTGAAGCTTTTGGTCTTCGCGGACCAGGAGAC CTCCAGGGAACTTTGGTGAT CTTCAATTGAATAAACTCGGCACTGAGGACCCACG TTGGCCCCAAATGCTGAGCTTGCT CCTACAGCCAGTGTCTTTATGGGTATGTCG CAATTTAACTTACCCATCAGAACATGAT GATCATGGCAACCCTGTGTACTTCTTCC GGTACAGTGGAGCCATTAACCTGACCCAAAG AATCCCAACTACAATAAGTGGTTGG AGCTTCTTGAGCAAAATATTGATGCCTACAAAACC TTCCCTAAGAAGGAAAAGAA CAAAAGGCACCAAAGAAGAAATCAACAGACCAAATGTCT GAACCTCAAAGGAGC AGCTGTGCAAGGTAGCATCACTCAGCGCACTCGCACCCGTCCA AGTGTTCA GCCTGGTCCAATGATTGATGTTAACACTGATTAG</p>
NCBI No.	NC_019843.3
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	02.25.2021
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*1vial	1ml/vial*1vial	√
Copy Number	>1x10 ⁷ copies/ml	2.27*10 ⁷ /ml	√
DNA Residue	<1%	0.87%	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****