

CERTIFICATE OF ANALYSIS

MERS-N Pseudovirus Standard Reference

Product information

Name	MERS-N Pseudovirus Standard Reference
Catalog No.	CBV30009
Lot No.	CGD2022022503
Format	Virus
Gene	Gene N
Sequence	<p>1242bp:</p> <p>ATGGCATCCCCCTGCTGCACCTCGTGCTGTTTCCTTTGCCGATAACAATGATATAACAAAT ACAAACCTATCTCGAGGTAGAGGACGTAATCCAAAACCACGAGCTGCACCAAATAACACT GTCTCTTGGTACACTGGGCTTACCCAACACGGGAAAGTCCCTTACCCTTCCACCTGGG CAGGGTGTACCTCTTAATGCCAATTCTACCCTGCGCAAAATGCTGGGTATTGGCGGAGA CAGGACAGAAAATTAATACCGGAATGGAATTAAGCAACTGGCTCCCAGGTGGTACTTC TACTACACTGGAAGTGGACCCGAAGCAGCACTCCCATTCCGGGCTGTTAAGGATGGC ATC GTTTGGGTCCATGAAGATGGCGCCACTGATGCTCCTTCAACTTTTGGGACGCG GAACCTT AACAATGATTCAGCTATTGTTACACAATTCGCGCCCGGTACTAAGCTTCT AAAAACTTC CACATTGAGGGGACTGGAGGCAATAGTCAATCATCTTCAAGAGCCTCT AGCTTAAGCAGA AACTCTTCCAGATCTAGTTACAAGTTCAAGATCAGGAACTTACC CGCGGCACCTTCT CCAGGTCCATCTGGAATCGGAGCAGTAGGAGGTGATCTACTTT ACCTTGATCTTCTGAAC AGACTACAAGCCCTTGAGTCTGGCAAAGTAAAGCAAT CGCAGCCAAAAGTAATCACTAAG AAAGATGCTGCTGCTGCTAAAAATAAGAT GCGCCACAAGCGCACTTCCACCAAAGTTTC AACATGGTGCAGCTTTTGGTCT TCGCGGACCAGGAGACCTCCAGGGAACTTTGGTGAT CTTCAATTGAATAAACT CGGCACTGAGGACCCACGTTGGCCCCAAATGCTGAGCTTGCT CCTACAGCCAGT GCTTTTATGGGTATGTCGCAATTTAACTTACCCATCAGAACATGAT GATCAT GGCACCCCTGTGTACTTCTTCCGTTACAGTGGAGCCATTAACCTGACCCAAAG AATCCCAACTACAATAAGTGGTTGGAGCTTCTTGGCAAAATATTGATGCCTAC AAAACC TTCCCTAAGAAGGAAAAGAAACAAAAGGCACCAAAGAAGAAATCAAC AGACCAAATGTCT GAACCTCAAAGGAGCAGCGTGTGCAAGGTAGCATCACT CAGCGCACTCGCACCCGTCCA AGTGTTCCAGCTGGTCCAATGATTGATGTTA ACACTGATTAG</p>
NCBI No.	NC_019843.3
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	02.25.2022
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*1vial	1ml/vial*1vial	√
Copy Number	>1x10 ⁷ copies/ml	3.57*10 ⁷ /ml	√
DNA Residue	<1%	0.92%	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****